



Rapid and simple As(III) quantification using a turbidimetric test for the monitoring of microbial arsenic bio-transformation

Caroline Michel, Catherine Joulain, Hafida Tris, Flavie Charnois, Fabienne Battaglia-Brunet

► To cite this version:

Caroline Michel, Catherine Joulain, Hafida Tris, Flavie Charnois, Fabienne Battaglia-Brunet. Rapid and simple As(III) quantification using a turbidimetric test for the monitoring of microbial arsenic bio-transformation. Journal of Microbiological Methods, 2020, 177, pp.106026. 10.1016/j.mimet.2020.106026 . hal-03517511

HAL Id: hal-03517511

<https://brgm.hal.science/hal-03517511>

Submitted on 30 Aug 2022

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.



Distributed under a Creative Commons Attribution - NonCommercial 4.0 International License

For submission to Journal of Microbiological Methods

Rapid and simple As(III) quantification using a turbidimetric test for the monitoring of microbial arsenic bio-transformation

Caroline MICHEL*, Catherine JOULIAN, Hafida TRIS, Flavie CHARNOIS, Fabienne BATTAGLIA-BRUNET

BRGM, DEPA (Direction de l'Eau, de l'Environnement, des Procédés et Analyses), 3 avenue Claude Guillemin, BP 36000, 45060 Orléans cedex 2, France

*Corresponding author: c.michel@brgm.fr

Abstract

A turbidimetric test for rapid quantification of As(III) (detection limit of 3 mg/L, quantification range of 10-100 mg/L) in liquid growth medium was developed for assessing and monitoring microbial As(III)-oxidizing and As(V)-reducing activities. This test is based on As(III) chelation with pyrrolidine dithiocarbamate followed by spectrometric measurement of absorbance, and was validated by comparison with AAS quantification of As after As(III)/As(V) separation.

Key words: arsenite quantification, pyrrolidine dithiocarbamate (PDTC), turbidimetric method, laboratory scale analysis, arsenic bio-transformation.

Arsenite (As(III)) is a major pollutant in many countries and is responsible of major sanitary problems due to the use of As-rich groundwater or surface water as drinking water. In view of the bioremediation of As(III)-polluted sites based on microbial As(III) oxidation into arsenate (As(V)) that precipitates more easily, many researches are directed toward the isolation and characterization of microbial strains and consortia able to oxidize As(III) (Santini et al., 2000 ; Battaglia-Brunet et al., 2002; Ike et al., 2008 ; Osborne et al., 2010 ; Paul et al., 2018). These works are generally screening for microorganisms highly tolerant to As and are usually conducted at high As(III) concentrations (until 100 mg/L and even more) and need analytical tools that allow total As quantification and also As(III) and As(V) species quantification at high concentrations. Standard analytical methods commonly used for arsenic (As) quantification (atomic absorption spectroscopy, mass spectroscopy) are sensitive with low detection limits and allow redox state determination. However, their main drawbacks are the equipment cost, the need of specific qualification of the experimenters, and the need to transport and store samples from the site to the laboratory. Several colorimetric tests have been developed to overcome these drawbacks (Dhar et al., 2004 ; Baghel et al., 2007 ; Hu et al., 2012 ; Sidhu et al., 2014; Sirawatcharin et al., 2014; Rathore, 2017 ; Kearns and Edson, 2018 ; Lace et al., 2019). They are cheap methods that allow routine and high frequency monitoring of polluted sites (that can be used as alert systems), and screening of samples that need to be further analyzed by standard analytical methods. However, as most of these colorimetric methods have been developed for on-site measurements of As in drinking waters, they usually display low detection limit and quantification ranges in the $\mu\text{g/L}$ order, and for many of them, redox state determination (As(III) vs. As(V)) is not possible. With the exception of few of them (Hu et al., 2012), most of these methods are not adapted to the characterization and monitoring of microorganisms in a context of site bioremediation. There is thus a need for rapid and simple laboratory tests that allow the quantification of As species, such as As(III), at high concentrations (until several dozens of mg/L).

The objective of this work was to develop a simple and rapid method to quantify As(III) (but not As(V)) in the 0-100 mg/L range in microbial culture media in the frame of growth, selection and

characterization of strains and consortia that express As(III)-oxidizing activity. In this context, the possibility to use the metals/metalloids chelation properties of pyrrolidine dithiocarbamate (PDTC) ($C_5H_9NS_2$) was evaluated. This chemical compound is known to form a non-soluble complex with As(III) and not with As(V) at pH 5, thus allowing to determine the redox state of As (Charlot, 1974; Subramanian and Meranger, 1981).

The procedure for the proposed PDTC test is simple and rapid (per sample: 5 minutes from sample preparation to absorbance measurement with a UV-visible spectrophotometer). In a glass tube, 0.5 mL of filtered growth medium (at 0.22 μm) are mixed with 0.5 mL of acetate buffer (0.1 M Na-acetate, pH 5 (pH is adjusted using acetic acid)). Then, 0.1 mL of a PDTC stock solution (5 g of PDTC (Sigma) dissolved in 1 L of demineralized water) are added and immediately mixed by a rapid hand shaking (by inverting the tubes 2-3 times). In the presence of As(III), white small precipitates immediately appear (whereas there are no precipitates with As(V)). Absorbance (at 450, 500, 550 and 600 nm) is then measured using a spectrophotometer (in this study: Cary 100 UV-Vis spectrophotometer, Agilent Technologies, using 1 cm light path cuvettes). Absorbance measurements have to be done immediately after shaking, as the precipitates tend to aggregate after 1 minute. It is thus recommend to analyze one sample at a time. The acetate buffer and PDTC stock solutions can be stored at room temperature for several months.

The first step of this work aimed to validate the possibility to quantify the turbidity due to As(III)-PDTC interaction using spectrophotometry. For this, the PDTC test was applied on As(III) solutions at various concentrations ranging from 0 to 100 mg/L and prepared in distilled water. Results showed a linear relation (R^2 about 0.98) between absorbance and As(III) concentration for the four tested wavelengths, and for As(III) concentrations ranging from 0 to 100 mg/L (Fig. 1). The calibration curves varied according to the wavelength so that the sensitivity was the highest at 450 nm and the weakest at 600 nm (Fig. 1, Table 1). The potential influence of microbial growth medium composition (as the presence of salts) on the sensitivity was then evaluated. For this, As(III) solutions prepared in CAsO1 medium (a

defined growth medium used for the growth of As(III)-oxidizing bacteria, Battaglia-Brunet et al., 2002) or 2A-MOPS medium (a defined growth medium adapted to marine As(III)-oxidizing bacteria, Lescure et al., 2013) were tested. Results demonstrated the influence of growth medium composition as sensitivity and quantification range varied according to the medium (Table 1). Sensitivity was indeed slightly higher for the CAsO1 medium in comparison to results obtained with distilled water and 2A-MOPS medium. This is potentially due to the presence, in the CAsO1 medium only, of Mg^{2+} and Ca^{2+} that probably favored the flocculation of precipitates (Sworska et al., 2000). Same sensitivity and detection range were obtained in the CAsO1 medium prepared with or without yeast extract (data not shown), showing no impact of yeast extract on the quantification of As(III). Concerning quantification range, for the CAsO1 medium, the linear relation between As(III) concentration and absorbance was lost for concentrations higher than 40 mg/L, at which a plateau was observed (data not shown) because of high absorbance values incompatible with the Beer-Lambert equation. To overcome this bias, if observed, a dilution of the sample, using growth medium as the dilution solution, is needed before analysis with the PDTC test. For the 2A-MOPS medium, As(III) quantification was possible until 100 mg/L range. More precisely, the detection limit was about 3 mg/L in water and 2A-MOPS medium, and about 2 mg/L in CAsO1 medium. However, linearity was optimal for concentrations higher than 10 mg/L. Taken all together, these results demonstrated the relevance of the PDTC test for quantifying As(III) in culture media containing up to 100 mg/L of As(III) (according to the growth media), and even more by diluting the sample. For accurate sensitivity of the method, the calibration curves have to be done in the growth medium used for the studied microbial strain or consortium. Moreover, to have the best As(III) concentration determination, it is suggested to measure the absorbance at the 4 wavelengths to calculate an average As(III) concentration.





Fig 1: Calibration curves obtained for the quantification of As(III) in distilled water by the PDTc test. Absorbance measurements ($n=3$) were registered at four different wavelengths (450 nm, 500 nm, 550 nm and 600 nm). Dotted lines represent trend lines that correspond to calibration curves. Bars represent standard deviation.

Wavelength (nm)	Sensitivity for As(III) (Abs / (mg/L))				Quantification range (mg/L)
	450 nm	500 nm	550 nm	600 nm	
In Water	0.0192	0.0148	0.0118	0.0095	10 - 100
In CAsO1 medium	0.0323	0.0232	0.0174	0.0136	10 - 40
In 2A-MOPS medium	0.0182	0.0141	0.0112	0.0092	10 - 100

Table 1: Sensitivity and quantification range of the PDTc test for As(III) according to the medium (CAsO1 medium, 2A-MOPS medium and distilled water) and to the wavelength used for absorbance measurement. Sensitivity values were calculated from trend lines such as those presented for water in Fig 1.

The PDTC test was then validated by comparison with a standard analytical method for As quantification and routinely applied for the monitoring of As(III)-oxidizing or As(V)-reducing activity tests (Lescure et al., 2016 ; Thouin et al., 2016). For this, As(III) was selectively extracted in Methyl IsoButyl Ketone (MIBK) with PDTC, using a protocol adapted from Charlot (1974) and detailed in Battaglia-Brunet et al. (2002). Total As and As(V) recovered in the aqueous phase were analyzed by Atomic Absorption Spectrophotometry (AAS; Varian AA22FS). The detection limit of this method was 1 mg/L. Standard As(III) solutions from 0 to 100 mg/L were prepared in distilled water, CASO1 medium and 2A-MOPS medium. As(V) and total As were quantified by AAS and As(III) was quantified by the PDTC test. For quantification using the PDTC test, calibration curves obtained in Figure 1 and Table 1 were used for the calculation of As(III) concentrations from absorbance values. For the two growth media and water, results from AAS and PDTC test showed similar results allowing the validation of the PDCT test (Fig. 2).






Fig 2 : Comparison of theoretical As(III) concentrations and measured As(III) concentrations by the PDTC test (squares, n=3) and AAS (circles, n=1). For the AAS approach, As(III) concentrations were obtained by deduction from total As and As(V) concentrations. Bars represent standard deviation.

Finally, the use of the PDTC test to measure and monitor As(III) oxidation by a bacterial strain was evaluated. For this, the As(III)-oxidizing bacterium *Thiomonas delicata* subsp. *arsenivorans* (Battaglia-Brunet et al., 2011) was grown in the CAsO1 medium at 25°C, in static conditions. Its growth was monitored by microscopic cell counts (using a Thoma cell (Marienfeld, depth 0.1 mm), and by examining 10 fields with a Zeiss Axio Imager Z1, magnification x400, AxioVision 4.6 software), and its As(III)-oxidizing activity was monitored by measuring As(III) concentration in the culture medium using the PDTC and AAS (for validation) approaches (Fig. 3). Abiotic tests (no inoculation) confirmed the stability of As(III) at 25°C in the CAsO1 medium over time (data not shown). Results showed that the As(III) oxidation kinetics was the same whatever the analytical method used for As(III) quantification, thus confirming the possibility to use the PDTC test for monitoring the activity of microbial strains able to oxidize As(III). Although not tested here, the developed test is also adapted to monitor the activity of As(V)-reducing microorganisms, since As(III) is the end product during the biological reduction of As(V).




Fig 3 : Growth and As(III)-oxidizing activity of *Thiomonas delicata* subsp. *arsenivorans* grown in the CASO1 medium. As(III) was quantified using the PDTC test (n=3) (results are means of values obtained at 550 nm and 600 nm), and AAS was used to quantify total As and As(V), thus allowing by deduction to obtain As(III) concentration (n=3). Bacterial growth was monitored by bacteria counting using a Thoma cell. Bars represent standard deviation.

In conclusion, a rapid (5 minutes per sample), simple (one reagent and one step) and cost effective (reagent (PDTC) cost is about 60 € for the preparation of a 5 L solution) turbidimetric method was developed for the quantification of As(III) (but not As(V)) in the range of 0 to 100 mg/L (according to the growth medium) or even more due to the possibility to dilute samples. This kind of test is particularly adapted to the characterization of As(III)-oxidizing microbial strains or consortia, but also

to As(V)-reducing microorganisms, as it allows to easily quantify and monitor their As-linked activity. This test, used here at the laboratory scale, could also be helpful at the (semi)-industrial scale for the development and monitoring of bioprocesses (prototypes or industrial installations) for bioremediation purposes such as treatment of As polluted water. Indeed, easy, cheap and rapid tests are needed at all scales to monitor bioremediation process efficiency, alert if pics of As(III) pollution are observed, and also select samples of interest that need further characterization using standard protocols. A third application would be the easy and cheap determination of microbial As(III)-oxidizing and As(V)-reducing activities in environmental samples, that proved to be interesting biogeochemical indicators (Lescure et al., 2016; Thouin et al., 2016). At last, preliminary tests realized in our laboratory also suggested that the developed PDTC turbidimetric test could be used to quantify Sb(III) (but with a lower sensitivity than that obtained for As(III)), another metalloid oxidized by some bacteria (Wang et al., 2015). PDTC was also demonstrated to interact with selenium (Subramanian and Meranger, 1981), suggesting that the PDTC test could potentially be used for the characterization of the bioremediation properties of microorganisms involved in the arsenic, antimony and/or selenium cycles.

This work was funded in the frame of de GEOMIC project (BRGM). The authors would like to thank Cindy Gassaud and Dominique Breeze for their involvement in laboratory work.

References

- Baghel, A., Singh, B., Pandey, P., Sekhar, K., 2007. A rapid field detection method for arsenic in drinking water. *Anal. Sci.* 23, 135-137.
- Battaglia-Brunet, F., Dictor, M.-C., Garrido, F., Crouzet, C., Morin, D., Dekeyser, K., Clarens, M., Baranger, P., 2002. An arsenic(III)-oxidizing bacterial population: selection, characterization, and performance in reactors. *J. Appl. Microbiol.* 93, 656–667.

180 Battaglia-Brunet, F., El Achbouni, H., Quemeneur, M., Hallberg, K. B., Kelly, D. P., Joulain, C., 2011.
 181 Proposal that the arsenite-oxidizing organisms *Thiomonas cuprina* and '*Thiomonas arsenivorans*' be
 182 reclassified as strains of *Thiomonas delicata*, and emended description of *Thiomonas delicata*. Int. J.
 183 Syst. Evol. Microbiol. 61, 2816–2821.

184 Charlot, G., 1974. Chimie analytique quantitative. Part II: Méthodes sélectionnées d'analyse chimique
 185 des éléments ed. Masson et Cie pp. 12 338–340.

186 Dhar, R., Zheng, Y., Rubenstone, J., van Geen, A., 2004. A rapid colorimetric method for measuring
 187 arsenic concentrations in groundwater. Analytic. Chim. Acta 526, 203-209.

188 Hu, S., Lu, J., Jing, C., 2012. A novel colorimetric method for field arsenic speciation analysis. J. Environ.
 189 Sci. 24, 1341-1346.

190 Ike, M., Miyazaki, T., Yamamoto, N., Sei, K., Soda, S., 2008. Removal of arsenic from groundwater by
 191 arsenite-oxidizing bacteria. Water Sci. Technol. 58, 1095-1100.

192 Kearns, J., Edson, C., 2018. Expanding quantification of arsenic in water to 0 µg L⁻¹ with a field test kit:
 193 substituting 0.4% M/V silver nitrate as the colorimetric reagent; employing digital image analysis.
 194 Water Air Soil Pollut. 229, doi.org/10.1007/s11270-018-3717-1.

195 Lace, A., Ryan, D., Bowlett, M., Cleary, J., 2019. Arsenic monitoring in water by colorimetry using an
 196 optimized leucomalachite green method. Molecules 24, doi :10.3390/molecules2402339.

197 Lescure, T., Carpentier, A., Battaglia-Brunet, F., Morel-Desrosiers, N., 2013. Oxidation of As(III) by the
 198 bacterial community of a marine sediment monitored by microcalorimetry. Geomicrobiol. J. 30, 540-
 199 548.

200 Lescure, T., Moreau, J., Charles, C., Ben Ali Saanda, T., Thouin, H., Pillas, N., Bauda, P., Lamy, I.,
 201 Battaglia-Brunet, F. (2016) Influence of organic matters on AsIII oxidation by the microflora of polluted
 202 soils. Environ. Geochem. Health 38, 911–925, DOI 10.1007/s10653-015-9771-3.

203 Osborne, T., Jamieson, H., Hudson-Edwards, K., Nordstrom, D., Walker, S., Ward S, Santini, J., 2010.
 204 Microbial oxidation of arsenite in a subarctic environment: diversity of arsenite oxidase genes and
 205 identification of a psychrotolerant arsenite oxidiser. BMC Microbiol. 10, 205.

206 Paul, T., Chakraborty, A., Islam, E., Mukherjee, S., 2018. Arsenic bioremediation potential of arsenite-
 207 oxidising *Micrococcus* sp. KUMAs15 isolated from contaminated soil. Pedosphere 28, 299-310.

208 Santini, J., Sly, L., Schnagl, R., Macy, A., 2000. A new chemolithoautotrophic arsenite-oxidizing
 209 bacterium isolated from a gold mine: phylogenetic, physiological, and preliminary biochemical studies.
 210 Appl. Environ. Microbiol. 66, 92-97.

211 Rathore, D., 2017. Quantifying arsenic metal in drinking water. Comments on: Highly sensitive & low
 212 cost colorimetric method for quantifying arsenic metal in drinking water of Malwa Punjab and
 213 comparison with ICAP-AES. J. Chromatogr. Sep. Tech. 8, DOI : 10.4172/2157-7064.1000386.

214 Sworska, A., Laskowski, J.S., Cymerman, G., 2000. Flocculation of the Syncrude fine tailings. Part I.
 215 Effect of pH, polymer dosage and Mg^{2+} and Ca^{2+} cations. Int. J. Miner. Process. 60, 143–152.

216 Sidhu, M., Mahajan, P., Bhatt, S., 2014. Highly sensitive & low cost colorimetric method for quantifying
 217 arsenic metal in drinking water of Malwa punjab and comparison with ICAP-AES. Annals Biol. Res. 5,
 218 105-109.

219 Sirawatcharin, S., Saithongdee, A., Chaicham, A., Tomapatanaget, B., Imyim, A., Praphairaksiti, N.,
 220 2014. Naked-eye and colorimetric detection of Arsenic (III) using difluoroboron-curcumin in aqueous
 221 and resin bead support systems. Analytical Sci. 30, 1129-1134.

222 Subramanian, K., Meranger, J., 1981. Determination of arsenic(III), arsenic(V), antimony(III),
 223 antimony(V), selenium(IV) and selenium (VI) by extraction with ammonium
 224 pyrrolodinedithiocarbamate-methyl isobutyl ketone and electrothermal atomic absorption
 225 spectrometry. Analytical Chim. Acta 124, 131-142.

226 Thouin, H., Le Forestier, L., Gautret, P., Hube, D., Laperche, V., Dupraz, S., Battaglia-Brunet, F., 2016.
227 Characterization and mobility of arsenic and heavy metals in soils polluted by the destruction of
228 arsenic-containing shells from the Great War. *Sci. Tot. Environ.* 550, 658–669.

229 Wang, Q., Warelow, T.P., Kang, Y.-S., Romano, C., Osborne, T.H., Lehr, C.R., Bothner, B., Mc Dermott,
230 T.R., Santini, J.M., Wanga, G., 2015. Arsenite oxidase also Functions as an antimonite oxidase. *Appl.*
231 *Environ. Microbiol.* 81, 1959-1965.

232